



AGENCIJA REPUBLIKE SRBIJE
ZA VEŠTAČENJE I ISTRAŽIVANJE
U OBLASTI FARMACEVTSKIH I
KEMIJSKIH ANALIZA
IZ OBLASTI FARMACEVTSKIH I
KEMIJSKIH ANALIZA

EXPERT RESEARCH PROTOCOL

from 08/22/2015

Code: 15-08-20-1 (103)

Customer: XXXXXXXXXXXXXXXXXXXXXXXXXXXX

Number of samples: Qualitative analysis of the sample, determination of purity

Methods: Agilent 1200, High-performance liquid chromatography (HPLC);
Column: Zorbax SB-C18 150 mm×2.1 mm, 3 mkm;
Detector – DAD, wavelength – 290, 245, 260 nm;
Detector – MSD, ionization method APCI Positive/Negative, SCAN (100-500 m/z)

Number of samples: 1

Subject: Anadrol (Oxymetholone)

Mobile phase: A - ACN-1% Formic acid (92%), B - H₂O-1% Formic acid (8%).

The elution mode is isocratic. The flow rate through the column: 0.3 ml/min. Thermostat: 30°C.

Single quadrupole mass analyzer is used for identification of the chemical elements. The samples were ionized at the electrostatic spraying (ESI) and atmospheric pressure with chemical ionization (APCI) mode with fixed positive and negative ions.




Fig.1. The component output chromatogram of the sample, detector DA, MS in EP, MS in AP

Tbl. 1. The calculation results of the peak areas on Fig. 1

Area Percent Report

Sorted By : Signal
Multiplier : 1.0000
Dilution : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: DAD1 A, Sig=290,4 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	2.547	BB	0.0840	5100.99902	929.99921	100.0000
Totals :				5100.99902	929.99921	

Signal 2: MSD1 TIC, MS File

Peak #	RetTime [min]	Type	Width [min]	Area	Height	Area %
1	2.623	BB	0.1475	6.65338e6	6.66708e5	100.0000
Totals :				6.65338e6	6.66708e5	

Signal 3: MSD2 TIC, MS File

Peak #	RetTime [min]	Type	Width [min]	Area	Height	Area %
1	2.615	BB	0.1124	1.88831e5	2.71058e4	100.0000
Totals :				1.88831e5	2.71058e4	

The analysis results of the received peaks by detector DA are shown in Fig. 2




Fig. 2. The analysis of the peak 1, DA detector

The analysis results of the received peaks by detector MS are shown on Fig. 3




Fig. 3. The analysis of the peak 1, MS detector (modes AP and AN)

The received data of the analytical comparison results of MS detectors with the calculated data on the test substance allow us to state that peak 1 refers to Oxymetholone (mw 332.4).

The chromatographic purity of Oxymetholone by MS and DA detectors is 98%.

Remarks:

1. This document was issued based on the results of examination of provided samples. This document does not guarantee the overall quality of the products.
2. This document can not be used for advertising purposes. It can not serve as an evidence in a lawsuit or be used for other purposes.
3. This document is valid for three months from the date of issue. Control samples (tests) are stored for one month and then disposed of.